




RevoFit™
by CLICK
Replacement Lace


PK0060-000-10



Replacement reel for all RevoFit™ kits

For help and video instructions, visit clickmedical.co/rfkb

Kit Contents:



- 1 x Replacement Lace
- 1 x Lace Feeder
- 1 x T6 Tool



WARRANTY:

www.clickmedical.co/terms/#warranty

U.S. Customers:

Contact Click Medical directly
Help@ClickMedical.co • Help Line: +1 970 670 7012

International Customers:

Please contact your local distributor



MDSS GmbH
Schiffgraben 41
30175 Hannover,
Germany




Click Medical, LLC
1205 Hilltop Parkway, W101
Steamboat Springs, CO,
80487, USA
+1-970-670-7012



Revised 2019

1 Remove reel.

- a. Pull the dial open.
- b. Use T6 tool to press inward on tab.
- c. Simultaneously turn reel housing clockwise 1/4 turn.





Release tab location




Always between tube path

If socket lacks a pilot hole, use vise grips to remove reel.


- 1 Locate tab (see location illustration above). Lock onto reel housing with needle-nose vise grips.
- 2 Rotate reel 5mm, then pull outward to remove. Cut one lace and pull to remove.



2 Remove old lace.




3 Clear tube holes.



Check tube holes for debris.
Check for rough edges and sand if rough to the touch.

4 Feed new lace through socket.





5 Feed lace through spool.




Feed lace into the small hole and out the big hole.
The knot lives in the big hole.

6 Tie knots.




Leave 8cm of lace on both ends before tying knots.



7 Install reel.

Insert reel fully into mounting plate, then rotate counter-clockwise to lock into place.



Pull panel out to keep lace from becoming pinched

Troubleshooting RevoFit™

Reduce the need to replace reels and laces by optimizing pad thickness and checking for sharp edges or debris.

Optimize Padding

Use pad thickness, material, and shape to refine pressure and fit.

Pad Material:


Medium density pads (shore: 35)

Pad Thickness:

Average 3/16" (4.75mm)


Pad Shape:

Slight convex shape allows for better application of pressure.






Check for correct pad thickness

CORRECT - Pad is the correct thickness when panel sits above socket when fit is snug to patient.



INCORRECT - Pad is too thin when panel is flush with socket.



Check for sharp edges and debris in the system.




Rinse system. This product is waterproof and submersible. Rinse with fresh water after use in saltwater, sand, or mud.




Clear tube holes. Check reel housing for excess tubing or debris.

1. Trim tube ends.
2. Clear all debris.



Check for rough edges.



Sand exit holes by hand until extra smooth.

Advisory and Weight Limit



Do not use on patients with neuropathy or circulatory issues.

Fabrication only by a certified professional trained in the RevoFit™ system.

Filament will wear; inspect and replace regularly.

Patient weight limit is 100kg per RevoFit™ system installed on the socket.

See advisories in other languages: clickmedical.co/advisories